

# BIOCOMPATIBILITY CERTIFICATE

**Material tested:** ELASTOSIL® LR 3040/50

**Manufacturer:** WACKER Chemie AG, Germany  
WACKER Chemical Corporation, USA  
WACKER Asahikasei Silicone Co. Ltd., Japan

**Studies performed:** The following studies were performed in order to determine the biocompatibility of the test item according to ISO 10993-1:

**CYTOTOXICITY** (BSL study no. 120843)

**SENSITIZATION** <sup>1)</sup> (BSL study no. 120844)

**TEST FOR PYROGENS** (BSL study no. 120845)

**Results:** The product did not show any adverse effects in the studies performed.

<sup>1)</sup> LLNA

**BSL BIOSERVICE Scientific Laboratories GmbH**

Behringstraße 6/8

D-82152 Planegg



Dr. Daniela Stelter  
Biological Safety Testing  
Date: 23 March 2012

